

Public Health Unit Infection Prevention and Control Lapse Report	
Initial Report	
Premise/facility under investigation (name and address)	Greenfield Dental
Type of premise/facility: (E.g. clinic, personal services setting)	Dental Office
Date Board of Health became aware of IPAC lapse	August 16, 2021
Date of Initial Report posting	August 31, 2021
Date of Initial Report update(s) (if applicable)	
How the IPAC lapse was identified	The complaint was self-reported by the dental practice
Summary Description of the IPAC Lapse	The office self-reported that an error was made in IPAC instrument processing procedure resulting in an isolated incident which resulted in an IPAC lapse. The complaint inspection revealed that the cleaning, disinfection and sterilization of medical devices on site does follow Provincial Infectious Disease Advisory Committee (PIDAC) Best Practice Standards for medical device reprocessing and infection control in a clinical office setting.
IPAC Lapse Investigation	
Did the IPAC lapse involve a member of a regulatory college?	Yes
If yes, was the issue referred to the regulatory college?	Yes
Were any corrective measures recommended and/or implemented?	The dental office is to ensure that: <ul style="list-style-type: none"> - Internal policies and procedures on infection control and medical device reprocessing are in place and followed. - Staff member(s) responsible for any or all steps in reprocessing must complete Public Health Ontario's (PHO) online learning in Reprocessing in Community Health Care Settings modules annually.
Please provide further details/steps	Some recommendations for practice improvements are in progress and require time to implement. Staff are committed to ensuring these improvements are made.
Date any order(s) or directive(s) were issued to the owners/operators (if applicable)	N/A
Initial Report Comments and Contact Information	
Any Additional Comments	

If you have any further questions, please contact:	
Name	Nicholas Dushenko
Title	Program Manager, Infectious Disease
E-mail Address	Nicholas.Dushenko@hnhss.ca
Phone number	519-426-6170 x. 3291
Final Report	
Date of final report posting	Sept 17 2021
Date any order(s) or directive(s) were issued to the owner/operator (if applicable)	
Brief description of corrective measures taken	Internal policies and procedures on infection control and medical device reprocessing are in place and followed.
Date all corrective measures were confirmed to have been completed	Sept 16 2021
Final report comments	The dental office is fully compliant with IPAC best standards of practice with office policy and in clinical practice.
Final Report Contact Information. If you have any further questions, please contact	
Name	Nicholas Dushenko
Title	Program Manager, Infectious Disease
E-mail Address	Nicholas.Dushenko@hnhss.ca
Phone number	519-426-6170 x. 3291